

CERTIFICATE OF ANALYSIS

PRODUCT PP1160P90
HE

LOT NUMBER 3424525

EXPIRY DATE 01.03.2022

DATE OF MANUFACTURE 04.01.2022

Delivery/Customer information

Date Printed
20.01.2022

Delivery No.
8711570632

Customer
511572
THERMO FISHER SCI. TAIWAN CO.
LTD
2101918

Physical Characteristics	Results	Specification
Appearance	Green to dark green coloured gel	Green to dark green coloured gel
pH	7,33	7,30 - 7,70

Sterility: Samples of this batch have been incubated at 35°C and room temperature to ensure product meets acceptable levels of sterility.

Microbiological Characteristics	Results	Specification Quantitative Tests
O2, 35°C, 18-24hr		
<i>Salmonella typhimurium</i> ATCC®14028	1-2mm blue green cols w/ black centres	
<i>Shigella flexneri</i> ATCC®12022	1-2mm blue green colonies	
<i>Escherichia coli</i> ATCC®25922	Marked inhibition; cols yellow to salmon	
<i>Enterococcus faecalis</i> ATCC®29212	Complete inhibition	

Released By



Nuriyana binti Nazri