

# CERTIFICATE OF ANALYSIS

**PRODUCT** PP1355P90  
XLD

**LOT NUMBER** 3427911

**EXPIRY DATE** 09.03.2022

**DATE OF MANUFACTURE** 12.01.2022

Delivery/Customer information
Date Printed 20.01.2022
Delivery No. 8711570643
Customer 511572 THERMO FISHER SCI. TAIWAN CO. LTD 2102127

Physical Characteristics	Results	Specification
Appearance	Red, slightly opalescent gel	Red, slightly opalescent gel
pH	7,35	7,20 - 7,60

**Sterility:** Samples of this batch have been incubated at 35°C and room temperature to ensure product meets acceptable levels of sterility.

Microbiological Characteristics	Results	Specification Quantitative Tests
O2, 35°C, 24hr		
<i>Salmonella typhimurium</i> ATCC®14028	1-2mm pink colonies with black centres	
<i>Shigella flexneri</i> ATCC®12022	1-2mm pink colonies	
<i>Escherichia coli</i> ATCC®25922	Marked inhibition (yellow colonies)	
<i>Enterococcus faecalis</i> ATCC®29212	Complete inhibition	

Released By



Noornabila binti Zainol